

**TEST CHANGE**

Urea Nitrogen, Fluid

0020183, FL UN

Specimen Requirements:

Patient Preparation:

Collect: Peritoneal/~~Ascites fluid.~~

Specimen Preparation: Centrifuge ~~and separate~~ to remove cellular material.  
Transfe~~port~~ 1 mL body fluid to an ARUP Standard Transport Tube. (Min: 0.2 mL)

Transport Temperature: Refrigerated.

Unacceptable Conditions: ~~Specimen~~Dialysate and any specimen types other than those listed.  
Specimens too viscous to be aspirated by instrument.

Remarks: Indicate source on test request form.

Stability: Ambient: ~~1 week~~24 hours; Refrigerated: ~~1 week~~5 days; Frozen: 1 year

Methodology: Quantitative Spectrophotometry

Performed: Sun-Sat

Reported: Within 24 hours

Note:

CPT Codes: 84520

New York DOH Approval Status: This test is New York DOH approved.

Interpretive Data:

For information on body fluid reference ranges and/or interpretive guidance visit <https://www.aruplab.com/bodyfluids> ~~Reference ranges for this assay have not been established for body fluid. Results should be interpreted in comparison to the concentration in blood and in conjunction with clinical context.~~

~~This test was developed and its performance characteristics determined by ARUP Laboratories. It has not been cleared or approved by the US Food and Drug Administration. This test was performed in a CLIA-certified laboratory and is intended for clinical purposes.~~

Reference Interval:

Reference Intervals have not been~~Not~~ established.

**HOTLINE NOTE:** There is a component change associated with this test. One or more components have been added or removed. Refer to the Hotline Test Mix for interface build information.