

TEST CHANGE

Extractable Nuclear Antigen Antibodies (SSA 52, SSA 60, and SSB)

0050791, SSA/SSB

Specimen Requirements:

Patient Preparation:

Collect: Serum separator tube or red tube.

Specimen Preparation: Separate serum from cells ASAP or within 2 hours of collection. Transfer 1 mL serum to an ARUP standard transport tube. (Min: 0.5 mL)

Transport Temperature: Refrigerated.

Unacceptable Conditions: Plasma or other body fluids. Contaminated, hemolyzed, or severely lipemic specimens.

Remarks:

Stability: After separation from cells: Ambient: 48 hours; Refrigerated: 2 weeks; Frozen: 1 year (avoid repeated freeze/thaw cycles)

Methodology: Semi-Quantitative Multiplex Bead Assay

Note:

CPT Codes: 86235 x3

New York DOH Approval Status: This test is New York DOH approved.

Interpretive Data:

Component	Interpretation
SSA-52 (Ro52) (ENA) Antibody, IgG	29 AU/mL or Less Negative 30-40 AU/mL Equivocal 41 AU/mL or greater Positive
SSA-60 (Ro60) (ENA) Antibody, IgG	29 AU/mL or Less Negative 30-40 AU/mL Equivocal 41 AU/mL or greater Positive
SSB (La) (ENA) Antibody, IgG	29 AU/mL or less Negative 30-40 AU/mL Equivocal 41 AU/mL or greater Positive

Reference Interval:

Test Number	Components	Reference Interval
	SSA-52 (Ro52) (ENA) Antibody, IgG	40 AU/mL or less
	SSA-60 (Ro60) (ENA) Antibody, IgG	40 AU/mL or less
	SSB (La) (ENA) Antibody, IgG	40 AU/mL or less

