

TEST CHANGE

Triiodothyronine, Reverse by Tandem Mass Spectrometry

2007918, RT3 TMS

Specimen Requirements:

Patient Preparation:

Collect: Plain red or serum separator tube. Also acceptable: Lavender (EDTA) or pink (K2EDTA).

Specimen Preparation: Allow serum specimen to clot completely at room temperature. Separate serum or plasma from cells ASAP or within two hours of collection. Transfer 2 mL serum or plasma to an ARUP Standard Transport Tube. (Min: 1 mL)

Transport Temperature: Frozen.

Unacceptable Conditions: Grossly hemolyzed specimens

Remarks:

Stability: Ambient: 24 hours; Refrigerated: 1 week; Frozen: 3 months

Methodology: Quantitative Liquid Chromatography-Tandem Mass Spectrometry

Performed: Sun-Sat

Reported: 1-54 days

Note:

CPT Codes: 84482

New York DOH Approval Status: This test is New York DOH approved.

Interpretive Data:

This test was developed and its performance characteristics determined by ARUP Laboratories. It has not been cleared or approved by the US Food and Drug Administration. This test was performed in a CLIA certified laboratory and is intended for clinical purposes.

Reference Interval:

Age	Reference Interval
0 - 17 years	Not established
18 years and older	9.0 - 27.0 ng/dl



*A nonprofit enterprise of the University of Utah
and its Department of Pathology*

Effective Date: **February 20, 2024**