

New Test **3002482** **Anti-sp100 and anti-gp210 Antibodies, IgG** **SP100GP210**



Additional Technical Information

Methodology: Semi-Quantitative Enzyme-Linked Immunosorbent Assay
Performed: Wed
Reported: 1-8 days

Specimen Required: Collect: Serum Separator Tube (SST).
Specimen Preparation: Separate from cells ASAP or within 2 hours of collection. Transfer 0.5 mL serum to an ARUP Standard Transport Tube. (Min: 0.3 mL)
Storage/Transport Temperature: Refrigerated.
Unacceptable Conditions: Non-serum, heat-inactivated, contaminated, grossly icteric, severely lipemic, grossly hemolyzed specimens or inclusion of fibrin clot.
Stability (collection to initiation of testing): After separation from cells: Ambient: 48 hours; Refrigerated: 2 weeks; Frozen: 1 year (avoid repeated freeze/thaw cycles)

Reference Interval:

Test Number	Components	Reference Interval	
3002478	Anti-sp100 Antibodies, IgG	20.0 Units or less	Negative
		20.1-24.9 Units	Equivocal
		25.0 Units or greater	Positive
3002477	Anti-gp210 Antibodies, IgG	20.0 Units or less	Negative
		20.1-24.9 Units	Equivocal
		25.0 Units or greater	Positive

Interpretive Data:

Refer to report

CPT Code(s): 83516 x2

New York DOH Approved.

HOTLINE NOTE: Refer to the Test Mix Addendum for interface build information.