

TEST CHANGE

COVID-19 IgG (Spike), Semi-Quantitative by CIA

3003648, COV19G SQ

Specimen Requirements:

Patient Preparation:

Collect: Serum separator tube (SST). Also acceptable: lithium heparin.

Specimen Preparation: Separate serum or plasma from cells ASAP or within 2 hours of collection. Transfer 1 mL serum or plasma to an ARUP Standard Transport Tube. (Min: 0.5 mL)

Transport Temperature: Refrigerated.

Unacceptable Conditions: Specimens containing particulate material or otherwise obviously contaminated. Severely hemolyzed, heat-inactivated, severely icteric, or lipemic specimens. Postmortem specimens.

Remarks:

Stability: After separation from cells: Ambient: 72 hours; Refrigerated: 1 week; Frozen: 1 month (avoid repeated freeze/thaw cycles)

Methodology: Semi-Quantitative Chemiluminescent Immunoassay (CLIA)

Performed: Sun-Sat

Reported: Within 24 hours

Note:

CPT Codes: 86769

New York DOH Approval Status: This test is New York DOH approved.

Interpretive Data:

This test is for in vitro diagnostic use under the FDA Emergency Use Authorization (EUA). In compliance with this authorization, please visit <https://www.aruplab.com/infectious-disease/coronavirus/testing> for more information and to access the applicable fact sheets.

Reference Interval:

Less than 1.00 Index Value	Negative
Greater than or equal to 1.00 Index Value	Positive



*A nonprofit enterprise of the University of Utah
and its Department of Pathology*

Effective Date: **February 6, 2023**