

HOTLINE: Effective June 7, 2021

New Test **3003913** **Orthopedic Metals Panel (Chromium, Cobalt, Titanium)** **ORTHO PAN**

Methodology: Quantitative Inductively Coupled Plasma-Mass Spectrometry (ICP-MS)
Performed: Varies
Reported: 7-10 days

Specimen Required: Collect: Body fluid.

Specimen Preparation: Transfer 4 mL body fluid to an ARUP trace element-free transport tube. (ARUP supply #43116 available online through eSupply using ARUP Connect™ or contact ARUP Client Services at (800) 522-2787.) (Min: 2.2 mL)

Test is not performed at ARUP; separate specimens must be submitted when multiple tests are ordered.

Storage/Transport Temperature: Refrigerated. Also acceptable: Room temperature or frozen.

Stability (collection to initiation of testing): Ambient: Undetermined; Refrigerated: Undetermined; Frozen: Undetermined

Reference Interval: By report

CPT Code(s): 82495; 83018 x 2

New York DOH approval pending. Call for status update.

HOTLINE NOTE: Refer to the Test Mix Addendum for interface build information.