

**NEW TEST – Available Now**

[Click for Pricing](#)

**Genital Ulcer Disease Panel by PCR**

3005674, GUDP PCR

**Specimen Requirements:**

**Patient Preparation:**

**Collect:** Genital, anal, or rectal swabs with APTIMA Unisex Swab Specimen Collection kit (ARUP supply #28907) OR in Viral Transport Media (ARUP supply #12884) available online through eSupply using ARUP Connect, or contact ARUP Client Services at (800) 522-2787.

**Specimen Preparation:** APTIMA Swab: Place blue swab in Swab Specimen Transport Tube, break shaft off at score line, then recap tube. Swab in Viral Transport Media (VTM): Transfer swab to viral transport media.

**Transport Temperature:** Frozen

**Unacceptable Conditions:** Serum, plasma, ocular fluid, and CSF

**Remarks:** Specimen source required

**Stability:** Ambient: 3 days; Refrigerated: 1 month; Frozen: 1 month

**Methodology:** Qualitative Polymerase Chain Reaction

**Performed:** Tues, Thurs, Sat

**Reported:** 1-5 days

**Note:** This test detects and differentiates Herpes simplex virus 1, Herpes simplex virus 2, Treponema pallidum, Haemophilus ducreyi, and Chlamydia trachomatis L serovar. This test does not differentiate Chlamydia trachomatis L1-L3 serovars.

**CPT Codes:** 87491, 87529 x2, 87798 x2

**New York DOH Approval Status:** Specimens from New York clients will be sent out to a New York DOH approved laboratory, if possible.

**Interpretive Data:**

A negative result does not rule out the presence of PCR inhibitors in the patient specimen or test-specific nucleic acid in concentrations below the level of detection by this assay.

This test was developed and its performance characteristics determined by ARUP Laboratories. It has not been cleared or approved by the U.S. Food and Drug Administration. This test was performed in a CLIA-certified laboratory and is intended for clinical purposes.

Reference Interval:

---

**HOTLINE NOTE: Refer to the Hotline Test Mix for interface build information.**