

**NEW TEST**

[Click for Pricing](#)

**Toxocara Antibodies, IgG by ELISA**

3006066, TOXOCARA G

**Specimen Requirements:**

**Patient Preparation:**

**Collect:** Serum separator tube (SST) or plain red.

**Specimen Preparation:** Transfer 1 mL serum to an ARUP standard transport tube. (Min: 0.5 mL)

**Transport Temperature:** Preferred transport temp: Refrigerated. Also acceptable: Frozen

**Unacceptable Conditions:** Contaminated, heat-inactivated, grossly hemolyzed, or lipemic specimens.

**Remarks:**

**Stability:** After separation from cells: Ambient: 48 hours; Refrigerated: 2 weeks; Frozen: 1 month

**Methodology:** Semiquantitative Enzyme-Linked Immunosorbent Assay

**Performed:** Wed, Sat

**Reported:** 1-7 days

**Note:** N/A

**CPT Codes:** 86682

**New York DOH Approval Status:** This test is New York DOH approved.

**Interpretive Data:**

Component	Interpretation
Toxocara Antibodies, IgG by ELISA	< 9 U Negative: No significant level of Toxocara IgG antibodies detected. 9 - 11 U Equivocal: Recommend repeat testing in 2-4 weeks. >11 U Positive: IgG antibodies to Toxocara detected,

	<p>indicating current or past infection. False-positive results due to infections with other helminths are possible.</p>	
--	--	--

Reference Interval:

Test Number	Components	Reference Interval
	Toxocara Antibodies, IgG by ELISA	8 U or less

**HOTLINE NOTE: Refer to the Hotline Test Mix for interface build information.**