

NEW TEST – Available Now

[Click for Pricing](#)

Complement Factor-H Concentration, Plasma

3020698, FACTOR-H

Specimen Requirements:

Patient Preparation:

Collect: Pink (K2EDTA), tan (K2EDTA), royal blue (K2EDTA), or lavender (EDTA).

Specimen Preparation: Separate plasma from cells within 2 hours of collection. Transfer plasma (minimum 0.5 mL) to an ARUP standard transport tube and freeze immediately.

Transport Temperature: CRITICAL FROZEN. Separate specimens must be submitted when multiple tests are ordered.

Unacceptable Conditions: Nonfrozen specimens. Specimens exposed to repeated freeze/thaw cycles. Grossly hemolyzed, lipemic, and icteric specimens. Serum samples. Heparinized and lithium plasma samples.

Remarks:

Stability: Ambient: Unacceptable; Refrigerated: Unacceptable; Frozen: 30 days

Methodology: Quantitative Enzyme-Linked Immunosorbent Assay (ELISA)

Note:

CPT Codes: 86160

New York DOH Approval Status: This test is New York DOH approved.

Interpretive Data:

Reference Interval:

Test Number	Components	Reference Interval
	Complement Factor-H conc, plasma	156-325 ug/mL

HOTLINE NOTE: Refer to the Hotline Test Mix for interface build information.